



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: March 29, 2012 INITIALS: DJK

OWNER: WASHINGTON STATE PATROL

DISTRICT: 5 ADDRESS:

LOCATION: FM SHOP

COUNTY:

MAKE: HAENNI

TYPE PORT. WHEEL SCALE

ACCURACY CLASS: IIII

MODEL: 101

ID. NO: 33617

CAPACITY: 20000 lbs. SECTIONS:

MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL

PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	14950	-50				
CENTER	20000	20000	0				


Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration",
Certification number(s): 36293-W-02

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: Dan Lafreniere

Date: April 3, 2012

Print name: DAN LAFRENIERE



State of Washington

County of Thurston

Signed and sworn to (or affirmed before me on April 3, 2012)

Patricia A. Geppert
(Signature of notary public)

Patricia A. Geppert
Printed Name of Notary

My appointment expires Sept. 4, 2015